




Date: July 15, 2025
Product: OSOM® Flu SARS-CoV-2 Combo Test (P/N 1080)
Subject: Kit Expiry Date Extension

Dear Valued Customer,

SEKISUI Diagnostics' request to extend the shelf life of the OSOM Flu SARS-CoV-2 Combo Test (P/N 1080) from 12 months to a maximum of 20 months was approved by the FDA on July 2, 2025.

This letter is to notify you that the OSOM Flu SARS-CoV-2 Combo Test kits in your inventory may now have a new expiration date which is different than what you see printed on the kit box.

Please consult the table attached to this message to verify the new expiry date for your kit(s). If your kit lot number is not included in the table, please use the expiration date printed on the outer kit box. For your convenience, a date expiration label and kit stuffer containing a QR code linking to Sekisui Diagnostics website is provided with-your kit. The QR code will link you directly to a list of kits that have been extended.



Please do not hesitate to contact SEKISUI Diagnostics Technical Support with any questions or concerns at 800-332-1042 or via email at techservices@sekisuidiagnostics.com.

Sincerely,

David Morris, PhD

SEKISUI Diagnostics
One Wall Street
Burlington, MA 01803 USA
+1-781-652-7800
techservices@sekisui-dx.com

Associate Director, Product Marketing

OSOM Flu SARS-CoV-2 Combo Test Kit Lot Number	Original Expiry Date	Updated Expiry Date
241114	10/31/2024	12/31/2025
241256	12/31/2024	2/28/2026
241323	2/28/2025	4/30/2026
241357	8/31/2025	4/30/2026
241357A	8/31/2025	4/30/2026
241421	9/30/2025	5/31/2026
241421A	9/30/2025	5/31/2026
241421A1	9/30/2025	5/31/2026
241421A2	9/30/2025	5/31/2026
241457	10/31/2025	6/30/2026
241461	10/31/2025	6/30/2026
241639	12/31/2025	8/31/2026
251035	1/31/2026	9/30/2026

SEKISUI Diagnostics
One Wall Street
Burlington, MA 01803 USA
+1-781-652-7800
techservices@sekisui-dx.com

TB-GEN-0140, REV 3.0
Effective date: July 15, 2025